Clinical Application of Soft Tissue Lasers in Periodontics

¹Sumit Narang, ²Rochira Maytreyee, ³Neha Singh, ⁴Kshma Khare, ⁵Prince Jain

ABSTRACT

Laser is an acronym for light amplification by stimulated emission of radiation. The use of laser in dentistry dates way back to 1960s. Since then, its use has increased manyfold. The dental lasers cut tissue by a principle known as photothermal interaction or photothermal ablation. Lasers are effectively used in operculectomies, gingival depigmentation, gingivectomies, and many more procedures. Laser cannot only be used for surgical procedures but can also be used as an adjunct to scaling and root planing. The diode laser is used on the soft tissue side of the periodontal pocket to remove the inflamed soft tissue and reduce the pathogens. Lasers have been suggested as an adjunct and/or alternative to conventional techniques for various periodontal procedures and considered superior with respect to easy ablation, decontamination, and hemostasis along with less operative and postoperative pain.

Keywords: Implant, Laser, Nonsurgical periodontal therapy, Root modification.

How to cite this article: Narang S, Maytreyee R, Singh N, Khare K, Jain P. Clinical Application of Soft Tissue Lasers in Periodontics. J Orofac Res 2015;5(3):94-98.

Source of support: Nil

Conflict of interest: None

INTRODUCTION

Lasers have been a part of dentistry for over 25 years, but because of their bulky body and cost, they have been largely ignored. The history of lasers dates back to 1960s, when Maiman introduced us to the concept of "light amplification by stimulated emission of radiation" (laser).¹ The first application of a laser to dental tissue was reported by Goldman et al and Stern and Sognnaes, each article describing the effects of the ruby laser on enamel and dentin.² There are various lasers available that can be used in periodontics. Optical properties of periodontium, such as pigmentation,

¹Professor and Head, ²⁻⁵Postgradute Student

¹⁻⁵Department of Periodontology, People's College of Dental Sciences and Research Centre, Bhopal, Madhya Pradesh India

Corresponding Author: Sumit Narang, Professor and Head Department of Periodontology, People's College of Dental Sciences and Research Centre, Bhopal, Madhya Pradesh India, Phone: +919300299399, e-mail: drsumitnarang19@gmail.com

water content, mineral content, heat capacity, and latent heats of transformation can also determine the clinical application alongside specific wavelength, heat conduction and dissipation, and the amount of tissue congestion.¹

Lasers for hard tissues encourage efficient diagnosis of caries and improve the resistance of dental enamel to caries, laser etching of enamel, cavity preparations, photopolymerization of composite resin, and sterilization of the root canal system.³ Various advantageous characteristics, such as hemostatic effects, selective calculus ablation, or bactericidal effects against periodontopathic pathogens, might lead to improved treatment outcomes. Because of an excellent soft tissue ablation capacity, CO₂ lasers have been used successfully as an adjunctive tool to de-epithelialize the mucoperiosteal flap during traditional flap surgery. Diode and Nd:yttrium aluminum garnet (YAG) lasers were mainly used for laser-assisted subgingival curettage and disinfection of the periodontal pocket with various degrees of success.⁴

MECHANISM OF LASERS

Laser is a monochromatic light in visible or invisible range with three indistinctive features of collimation, coherency, and efficiency. Based on the quantum theory of physics given by Max Plank and atomic architecture by Niels Bohr, the concept of spontaneous emission was proposed. It defined as the process by which a light source, such as an atom, molecule, nanocrystal, or nucleus in an excited state undergoes a transition to the ground state and emits a photon. In a laser, atoms are kept in an excited state by "pumping" the laser, and some photons are inserted that cause some atoms to undergo stimulated emission, and the resulting photons cause other atoms to undergo stimulated emission, leading to a chain reaction that produces an intense, coherent, and easily focused light.5 This phenomenon does not occur anywhere in nature.

CLASSIFICATION OF LASERS

Lasers can be classified according its spectrum of light (Table 1), material used (Table 2), hardness, etc. They are also classified⁶ on the basis of hardness:

| Table 1: Classification based on light spectrum | | | | | | |
|---|--------------|--|--|--|--|--|
| Ultraviolet light | 100–400 nm | Not used in dentistry | | | | |
| Infrared light | 400–750 nm | Most commonly used in dentistry (argon and diagnodent laser) | | | | |
| Visible light | 750–10000 nm | Most dental lasers are in this spectrum | | | | |

| Table 2: | Classification | according | to | material used | ł |
|----------|----------------|-----------|----|----------------|----|
| TUDIO L. | olucomoulion | according | .0 | material about | ×. |

| Gas | Liquid | Solid |
|----------------|----------------------------|---|
| Carbon dioxide | Not so far in clinical use | Diodes, Nd:YAG, Er:YAG, Er:Cr:YSGG, Ho:YAG |
| | 0 | |

YAG: yttrium aluminum garnet; YSGG: yttrium-selenium-galliumgarnet

- Soft lasers
- Hard lasers

LASER APPLICATIONS IN PERIODONTAL THERAPY

The use of laser in periodontal treatment is more complex as periodontium consists of both hard and soft tissues, so the erbium group of lasers seems more beneficial for periodontal applications as it can be used for both hard and soft tissues.¹

Nonsurgical Periodontal Treatment

Lasers have solved the limitation of mechanical instrumentation as the beam can reach the deepest pockets and grooves. Periodontopathogens are deactivated in 50°C, according to many researches^{7,8} and coagulation, inflamed tissue removal is achieved by laser at 60°C.^{7,8} Also, lasers offer clear field views, which are advantageous over the conventional treatments. For bacterial reduction and coagulation, soft tissue lasers, such as argon, diode laser,⁹ and Nd:YAG¹⁰⁻¹² are a good choice for inflamed tissues and pigmented bacteria. Er:YAG lasers had been shown to remove subgingival calculus without a thermal change of the root surface, which is equivalent to that provided by scaling and root planing.^{5,13,14}

Photodynamic therapy is another important application of a low-power laser that decontaminates the periodontal pocket through activation of a photosensitizer agent, which potentiates the bactericidal effect of laser.⁵

Periodontal Surgery

Soft tissue lasers, viz., CO₂, Nd:YAG, diode, Er:YAG, and Er,Cr: yttrium-selenium-gallium-garnet (YSGG), are being widely used as a tool for gingival soft tissue procedures such as gingivectomy, gingivoplasty, epulis or benign tumor removal, frenectomy,² gingival depigmentation, crown lengthening surgeries, irradiation of aphthous ulcers, second-stage exposure of dental implants, coagulation of free gingival graft donor sites. A summary of different laser wavelengths and their periodontal application can be seen in Table 3.¹⁵

The purported advantages of lasers *vs* scalpel surgery include increased coagulation that yields a dry surgical field and better visualization; the ability to negotiate curvatures and folds within tissue contours; tissue surface sterilization and, therefore, reduction in bacteremia; decreased swelling, edema, and scarring; decreased pain; faster healing response; and increased patient acceptance.² Most laser excisional or incisional procedures are performed at 100°C, where intra- and

| Table 3: Laser application in periodontics | | | | | | |
|---|-----------------|--|--|--|--|--|
| Laser type | Wavelength (nm) | Contact | Clinical applications in periodontics | | | |
| Carbon dioxide laser | 10,600 | Beam focused at 1–2 mm from target surface | Soft tissue incision and ablation; subgingival curettage; biopsy; depigmentation, decontamination of implant | | | |
| Neodymium:yttrium-aluminum- garnet (Nd:YAG) laser | 1,064 | Surface contact required | Soft tissue incision and ablation; subgingival curettage; bacterial elimination | | | |
| Erbium:yttrium-aluminum-garnet (Er:YAG) laser | 2,940 | Surface contact required | Soft tissue incision and ablation; subgingival curettage; scaling; root conditioning; osteoplasty and ostectomy; degranulation and decontamination of implants, melanin and metal pigment removal | | | |
| Erbium,chromium:yttrium-selenium- gallium-garnet (Er,Cr:YSGG) laser | 2,780 | Surface contact required | Soft tissue incision and ablation; subgingival curettage; scaling of root surfaces; osteoplasty and ostectomy | | | |
| Indium-gallium-arsenide-phosphide (InGaAsP); Gallium-aluminum- arsenide (GaAlAs); Gallium- arsenide (GaAs) (diode) laser | 635 to 950 | Surface contact required | Soft tissue incision and ablation; subgingival curettage; disinfection, sulcular debridement, melanin pigment removal | | | |
| Argon laser | 488 and 514 | Noncontact or no contact mode | Soft tissue incision and ablation, subgingival curettage; bacterial elimination, putpotomy, root canal disinfection sulcular debridement caries removal, aphtus ulcer treatment, analgesia, melanin pigment removal, treatment of dentine hypersentitivity | | | |

Table 3: Laser application in periodontics

extracellular water vaporizes and leads to ablation or removal of tissue. Clinicians should have knowledge that if the tissue temperature exceeds 200°C during the procedure, then carbonization and irreversible tissue necrosis will take place.¹⁶

A key factor in determining how the laser will interact with the underlying tissues depends on their penetration depth and hence may possibly damage the underlying tissues by thermal effects. In CO₂, Er:YAG, and Er,Cr:YSGG lasers, laser light is absorbed in superficial layers and hence is advantageous. However, deeply penetrating Nd:YAG and diode lasers having greater thermal effects, leaving a thicker coagulation area on the treated surface similar to electrosurgical procedures.^{1,2,5}

Effect of most dental lasers on bone is detrimental for osseous surgery. Exposure of bone to heating at levels $\pm 47^{\circ}$ C is reported to induce cellular damage leading to osseous resorption, and temperature levels of $\pm 60^{\circ}$ C result in tissue necrosis.² But, surprisingly, there are exceptions like Er:YAG and Er,Cr:YSGG laser, which can be used for ostectomy and osteoplasty.⁵ However, lower cutting efficiency as compared with conventional instruments and lack of depth control are limitations of these lasers.¹⁷

Root Surface Modifications

Root surface modification using CO₂, Nd:YAG, Er:YAG, and diode laser has been studied by many authors with conflicting results and is related to the energy density and laser wavelength selection. *In vitro* studies have shown better fibroblast adherence with Er:YAG laser irradiation than mechanical scaling alone,^{18,19} whereas conflicting results were seen with Nd:YAG laser.^{20,21}

Implant Therapy

Most commonly, lasers have been used in the implant dentistry for soft tissue removal during second-stage implant exposure.¹ Various lasers have been used in second-stage implant therapy for uncovering the submerged implant, prior to placement of the healing abutment. They offer advantages of improved homeostasis, fine cutting surface, less postoperative discomfort, and favorable healing. Although Er:YAG lasers have been used to prepare fixture holes in bone during first-stage implant therapy, because of difficult and time-consuming mechanical debridement, emergence of bacterial resistance to antibiotics, lasers are now being proposed for treating peri-implantitis.⁵ Among the dental lasers, Er:YAG lasers at appropriate settings possess the best property for degranulation and implant surface decontamination, without causing surface changes in implant.^{5,22}

Advantages of laser treatment are greater hemostasis, bactericidal effect, and minimal wound contraction.²²⁻²⁴ Compared with the use of a conventional scalpel, there is reduced need for anesthesia; lasers can cut, ablate, and reshape the oral soft tissue more easily, with no or minimal bleeding and little pain as well as no or only a few sutures.⁶ Thus it reduces stress and fatigue for the operator and patients with greater comfort during and after surgery with high acceptance of treatment. It also produces less collateral thermal damage than will an electrocautery. Lasers are also beneficial in managing peri-implantitis cases. Bactericidal properties and delicate debridement of microscopic fins and recesses are advantageous over conventional hand instruments.

Unlike antibiotics, lasers render their bactericidal properties without common side effects such as bacterial resistance, drug interactions, and gastrointestinal complications, and can thus be used in childhood and pregnancy without limitations. It has been reported that soft tissue surgery with CO_2 laser reduces the surgery time to one-fourth. But there is loss of tactile sense when using soft tissue lasers, such as CO_2 , diode, and Er:YAG.⁴

On the other side, the use of lasers also has disadvantages in that all lasers require specialized training and attention to safety precautions. Laser irradiation can interact with tissues even in the noncontact mode, which means that laser beams may reach the patients' eyes and other tissues surrounding the target in the oral cavity. Clinicians should be careful to prevent inadvertent irradiation to these tissues, especially to the eyes.⁶ Thus safety issues, technical complexities, cost, and lack of evidence-based studies about therapeutic effects and efficiencies are drawbacks of laser treatment.¹ Also, no single laser can perform all desired dental applications.

PRECAUTIONS AND RISKS ASSOCIATED WITH CLINICAL USE OF LASERS: PRECAUTIONS BEFORE AND DURING IRRADIATION²⁵

- Use glasses for eye protection (patient, operator, and assistants).
- Prevent inadvertent irradiation (action in noncontact mode).
- Protect the patient's eyes, throat, and oral tissues outside the target site.
- Use wet gauze packs to avoid reflection from shiny metal surfaces.
- Ensure adequate high-speed evacuation to capture the laser plume.

Potential Risks

- Excessive tissue destruction by direct ablation and thermal side effects.
- Destruction of the attachment apparatus at the bottom of pockets.
- Excessive ablation of root surface and gingival tissue within periodontal pockets.
- Thermal injury to the root surface, gingival tissue, pulp, and bone tissue.

RECENT ADVANCES

Waterlase system is a revolutionary dental device that uses laser-energized water to cut or ablate soft and hard tissue and provide periodontists with the opportunity to perform more procedures in fewer appointments with less need for anesthesia, scalpels, and drill.⁵

Periowave is a photodynamic disinfection system that utilizes nontoxic dye (photosensitizer) which, in combination with low-intensity lasers, enable singlet oxygen molecules to destroy bacteria. After applying light-sensitive drug (photosensitizer), low-intensity laser is directed on the area treated with the drug, resulting in phototoxic reactions. Although the use of photosensitizers for complete suppression of the anaerobic periopathogens has been suggested, however, the same is not true for facultative anaerobes.⁵

CONCLUSION

As technology advances in the field of dentistry, whether laser or any other exciting technology, the options available to clinicians will continue to increase. Although the use of lasers in dentistry is relatively new, the future looks very bright. In summary, laser treatment is expected to serve as an alternative or adjunctive to conventional mechanical periodontal treatment.⁶ Based on this review of the literature, there is a great need to develop an evidence-based approach to the use of lasers for the treatment of chronic periodontitis. Simply put, there is insufficient evidence to suggest that any specific wavelength of laser is superior to the traditional modalities of therapy. Safety regarding the use of lasers is not optional. It should be the highest priority of all clinicians involved in the surgical procedure. There should be a concerted effort to better educate physicians, provide more educational support for perioperative staff members, and conduct additional research to improve equipment and ancillary supplies.²⁶ Prevention of accidents requires thorough knowledge of their causes and the application of measures to avoid them.¹⁵ Introduction of lasers in implant therapy and

newer laser technical modalities has revolutionized the periodontal treatment outcome with patient acceptance. However, patient risk and procedural cost must always be considered and fully understood before its application.⁵

REFERENCES

- 1. Seyyadi SA, Khashabi E, Falaki F. Laser application in periodontics. J Lasers Med Sci 2012 Winter;3(1):26-32.
- 2. Cobb CM. Lasers in periodontics: a review of literature. J Periodontol 2006 Apr;77(4):545-564.
- 3. Singh S, Gambhir RS, Kaur A, Singh G, Sharma S, Kakar H. Dental lasers: a review of safety essentials. J Lasers Med Sci 2012 Summer;3(3):91-96.
- Schwarz F, Aoki A, Becker J, Sculean A. Laser application in non-surgical periodontal therapy: a systematic review. J Clin Periodontol 2008 Sep;35(Suppl 8):29-44.
- 5. Bains VK, Gupta S, Bains R. Lasers in periodontics: an overview. J Oral Health Comm Dent 2010;4(Suppl):29-34.
- 6. Dang AB, Rallan NS. Role of lasers in periodontology: a review. Ann Dent Specialty 2013;1(1):8-12.
- 7. Russell AD. Lethal effects of heat on bacterial physiology and structure. Sci Prog 2003;86(Pt 1-2):115-137.
- Knappe V, Frank F, Rohde E. Principles of lasers and biophotonic effects. Photomed Laser Surg 2004 Oct;22(5): 411-417.
- 9. Moritz A, Schoop U, Goharkhay K, Schauer P, Doertbudak O, Wernisch J. Treatment of periodontal pockets with a diode laser. Lasers Surg Med 1998;22(5):302-311.
- Neil ME, Mellonig JT. Clinical efficacy of the Nd:YAG laser for combination periodontitis therapy. Pract Periodontics Aesthet Dent 1997 Aug;9(6 Suppl):1-5.
- 11. Liu CM, Hou LT, Wong MY, Lan WH. Comparison of Nd: YAG laser versus scaling and root planing in periodontal therapy. J Periodontol 1999 Nov;70(11):1276-1282.
- 12. Raffetto N, Gutierrez T. Lasers in periodontal therapy, a five years retrospective. J Calif Dental Hygiene Assoc 2001;16: 17-20.
- Akoi A, Ando Y, Watanabe H, Ishikawa I. In vitro studies on laser scaling of subgingival calculus with an erbium:YAG laser. J Periodontol 1994 Dec;65(12):1097-1106.
- Akoi A, Miura M, Akiyama F, Nakagawa N, Tanaka J, Oda S, Watanabe H, Ishikawa I. In vitro evaluation of Er:YAG laser scaling of subgingival calculus in comparison with ultrasonic scaling. J Periodontal Res 2000 Oct;35(5):266-277.
- 15. Brown IS. Current advances in the use of lasers in periodontal therapy: a LANAP case series. Clin Adv Periodontics 2013 May;3(2):96-104.
- Lomke MA. Clinical applications of dental lasers. Gen Dent 2009 Jan-Feb;57(1):47-59.
- Ishikawa I, Aoki A, Takasaki AA, Mizutani K, Sasaki KM, Izumi Y. Application of lasers in periodontics: true innovation or myth? Periodontol 2000 2009;50:90-126.
- Belal MH, Watanabe H, Ichinose S, Ishikawa I. Effect of Er:YAG laser combined with rhPDGF-BB on attachment of cultured fibroblasts to periodontally involved root surfaces. J Periodontol 2007 Jul;78(7):1329-1341.
- Feist IS, De Micheli G, Carneiro SR, Eduardo CP, Miyagi S, Marques MM. Adhesion and growth of cultured human gingival fibroblasts on periodontally involved root surfaces

treated by Er:YAG laser. J Priodontol 2003 Sep;74(9): 1368-1375.

- 20. Thomas D, Rapley J, Cobb C, Spencer P, Killoy W. Effects of Nd:YAG laser and combined treatments on in vitro fibroblast attachment to root surfaces. J Clin Periodontol 1994 Jan;21(1):38-44.
- 21. Trylovich DJ, Cobb CM, Pippin DJ, Spener P, Killoy WJ. The effects of the Nd:YAG laser on in vitro fibroblast attachment to endotoxin treated root surfaces. J Periodontol 1992 Jul;63(7):626-632.
- 22. Schwartz F, Rothamel D, Sculean A, Georg T, Scherbaum W, Becker J. Effects of an Er: YAG laser and the vector ultrasonic system on the biocompatibility of titanium implants in cultures of human osteoblast-like cells. Clin Oral Implants Res 2003 Dec;14(6):784-792.
- 23. Ando Y, Aoki A, Watanabe H, Ishikawa I. Bactericidal effect of erbium YAG laser on periodontopathic bacteria. Lasers Surg Med 1996;19(2):190-200.
- 24. Kreisler M, Al Haj H, d'Hoedt B. Temperature changes at the implant–bone interface during simulated surface decontamination with an Er:YAG laser. Int J Prosthodont 2002 Nov-Dec;15(6):582-587.
- Yamaguchi H, Kobayashi K, Osada R, Sakuraba E, Nomura T, Arai T, Nakamura J. Effects of irradiation of an erbium: YAG laser on root surfaces. J Periodontol 1997 Dec;68(12): 1151-1155.
- Cohen RE, Ammons WF. Lasers in periodontics. Report of Research, Science and Therapy Committee of the American Academy of Periodontology (revised by Rossman LA). J Periodontol 2002 Oct;73(10):1231-1239.